



Humidity and Temperature Probe HMP7
For high humidity, with probe warming

		Probe type	Probe cable	Sensor type	Filter type	Sensor purge	Baud rate	Parity, Data bit, Stop	Modbus address	Reserved	Installation accessory	Connection cable
	Order code	HMPX	HMPX	7						0		
1	Probe type	HMP7 for high humidity	7									
2	Probe cable	2m cable <i>10m cable</i>	E F									
3	Sensor type	Humicap R2 composite sensor, allows sensor purge Catalytic composite sensor, allows sensor purge		1 3								
4	Filter type	Sintered stainless steel filter Membrane SST filter for fuel cell applications			B E							
5	Sensor purge, default purge interval 24h 1) 2) Probe heat on, purge on. Dew point measurement only					2						
6	Baud rate 2) 19200 bps 9600 bps						A B					
7	Parity, Data bit, Stop bit 2) 8, N, 2 8, E, 1 8, O, 1							0 2 4				
8	Modbus address 2) 240 110 120 130								A B C D			
9	Reserved None									0		
10	Installation accessory None <i>Duct installation kit</i> <i>Cable gland with split seal; for sealing the probe from the cable</i> <i>Swagelok for NPT 1/2" thread</i> <i>Swagelok for ISO 3/8" thread</i> <i>Swagelok for ISO 1/2" thread</i> <i>Splash protector for warmed HMP7</i>										0 B C G H J K	
11	Connection cable None <i>1.5m with flying leads</i> <i>10m, with flying leads</i>											0 1 2

- 1) When probe heating is activated, dew point temperature, mixing ratio and water mass fraction are available.
For RH, T and other variables, compensation temperature has to be set through the Modbus interface
- 2) Factory pre-set, can be changed in the field with a service cable (P/N 242659)

Probe can be connected to INDIGO series of transmitters regardless of the output configuration.

Selections in bold are included in the prices of the basic versions.
Selections in italic are available at an extra price.

Example of order code with typical settings:

For use with INDIGO transmitters	HMPX	7	E	2	A	6	A	0	A	0	0	0
For use with Modbus RTU	HMPX	7	E	2	A	6	A	0	A	0	0	2