
### **MODEL - N2**

# **DIRECT MOUNT BLOCK AND BLEED (Enclosure Pipe to Flange) TWO VALVE MANIFOLD** 413 bar (6000 psi)







## Valve Colour Coding APPLICATION

#### **Using The 2-Valve Manifold**

In normal operation the "isolate" valve is open while the "vent" valve is closed. To remove the instrument, first close the "isolate" valve, then open the "vent" valve to relieve pressure upstream of the "isolate" valve.


#### **Calibration Options**

By connecting a calibration gauge to the vent port, it is possible to check the calibration of the instrument without removing it from the installation.

Also available in a range of other materials and options (See HOW TO ORDER Data Sheet)



### WEIGHT = 2.8 kg



#### **STEWART-BUCHANAN GAUGES LTD**